
Authorized Distributor





Glimepiride and Telmisartan RSolv® Capella EX-C18, 5µm, 150 x 4.6 mm



Test Condition

Column: RSolv® Capella EX-C18, 5µm, 150 x 4.6 mm

Flow Rate: 0.6 mL/min

Injection volume: 10 µl

Detection: PDA 254 nm

Mobile Phase: 20 mM Dibasic Ammonium Phosphate: Acetonitrile (30:70 v/v)

Column Temperature: 35°C

Sample Temperature: 15°C

Discussion

The isocratic mobile phase consists of 1.32g dibasic ammonium phosphate in 750 mL of water, pH 2.5 with 5% v/v phosphoric acid, and dilute in 1000 mL water. From prepared buffer take 300 mL and add 700 mL acetonitrile (30:70 v/v) mix, and pass through a filter of 0.45 μ m pore size. A RSolv Capella C18, 150 x 4.6 mm, 5 μ m analytical column from RSolv, Made in USA, was used as a stationary phase. A constant flow rate of 0.6 mL/min was employed throughout the analysis. A variable PDA detector was set at 254 nm.

The chromatographic analysis detected a significant peak at 5.404 and 8.486 minutes, representing Glimepiride and Telmisartan with a tailing factor of 1.27, indicating a symmetrical peak. Theoretical plates for Glimepiride and Telmisartan were calculated at 13906 and 11877, indicating good chromatographic efficiency

For more information, Contact us at