



RIKI GLOBAL
performance through innovation

www.rikiglobal.com



Plot No 39 1st Floor Lavanya Aracade Jayabheri Enclave
Gachibowli Hyderabad-500032.Telangana
Mob : +91 99082 46685 Tel : 040 - 35172288/89

info@rikiglobal.com

www.rikiglobal.com

+91-9908246695 +91-40-35172288/89



RIKI GLOBAL
performance through innovation

ROMIL-SpA™ Super Purity Acids and Reagents Specifications

Acetic Acid glacial (see Acetic Acid)

Ammonium Hydroxide solution (see Ammonia solution)

Acetic Acid SpA

H015

500ml H015P
Dgr H:226-314
P:280c-301+330+331-305+351+338-
307+310



(Acetic Acid glacial)
CH₃COOH MW60.05 FP 16.7°C BP 117.9°C d 1.05 CAS [64-19-7]

Assay >99%
Colour <10 Hazen (APHA)
Be, Bi, Ce, Co, Cs, Dy, Er, Eu, Ga, Gd, Ho, In, La, Li, Lu, Nd, Pb, Pr, Rb, Re, Sc,
Sm, Tb, Th, Tl, Tm, U, Y, Yb, Zr <0.1 ppb each
As, Ba, Cd, Cu, Ge, Mg, Mn, Mo, Ni, Pt, Rh, Ru, Sb, Sn, Sr, Te, Ti, V, W <0.5 ppb
each
Ag, Al, Ca, Cr, Fe, Hg, K, Na, Se, Zn <1 ppb each
SO₄ <0.5 ppm
PO₄ <1 ppm
Cl <1 ppm
Substances reducing dichromate passes test
Substances reducing permanganate passes test
Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis (for organic trace analysis see ROMIL Code H014 Acetic Acid SpS), Ion Chromatography, Voltammetry

Elemental impurities specified at time of manufacture.

Typical values, elemental impurities at time of manufacture (ppb):
Ag <0.1 Cu <0.2 K <0.5 Pt <0.1 Te <0.1 Al <0.5 Dy <0.1 La <0.1 Rb <0.1 Th
<0.1 As <0.1 Er <0.1 Li <0.1 Re <0.1 Ba <0.1 Eu <0.1 Lu <0.1 Rh
<0.1 Tl <0.1 Be <0.1 Fe <0.5 Mg <0.2 Ru <0.1 Tm <0.1 Bi <0.1 Ga <0.1 Mn
<0.1 Sb <0.1 U <0.1 Ca <0.1 Gd <0.1 Mo <0.1 Sc <0.1 V <0.1 Cd <0.1 Ge
<0.1 Na <1 Se <0.5 W <0.1 Ce <0.1 Hf <0.1 Nd <0.1 Sm <0.1 Y <0.1 Co
<0.1 Hg <0.1 Ni <0.1 Sn <0.1 Yb <0.1 Cr <0.1 Ho <0.1 Pb <0.1 Sr <0.1 Zn
<0.5 Cs <0.1 In <0.1 Pr <0.1 Tb <0.1 Zr <0.1

Ammonia solution SpA

H058

500ml H058P
Dgr H:314-335
P:280c-301+330+331-304+340-
305+351+338-309+310



(Ammonium Hydroxide solution)
NH₃ MW 17.03 d 0.92CAS[1336-21-6]

Assay 20-22%
Colour <10 Hazen (APHA)
Ba, Be, Bi, Ce, Cs, Dy, Er, Eu, Ga, Gd, Ge, Ho, In, La, Li, Lu, Nb, Nd, Pb, Pr, Rb, Sc,
Sm, Sr, Tb, Te, Th, Tl, Tm, U, W, Y, Yb, Zr <0.1 ppb each
Hg <0.2 ppb
Ag, Au, Cd, Co, Cr, Cu, Mn, Mo, Ni, Rh, Sb, Sn, Ti, V, Zn <0.5 ppb each
Al, As, Ca, Fe, K, Mg, Na, Se <1 ppb each
PO₄ <0.01 ppm
Cl <0.5 ppm
SO₄ <1 ppm
Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis, Ion Chromatography, Voltammetry



Elemental impurities specified at time of manufacture.

Typical values, elemental impurities at time of manufacture (ppb):
Ag <0.1 Cu <0.5 La <0.1 Pt <0.1 Th <0.1 Al <0.5 Dy <0.1 Li <0.1 Rb <0.1 Ti
<0.1 As <0.1 Er <0.1 Lu <0.1 Re <0.1 Tl <0.1 Au <0.1 Eu <0.1 Mg <0.2 Rh
<0.1 Tm <0.1 Ba <0.1 Fe <0.5 Mn <0.2 Ru <0.1 U <0.1 Be <0.1 Ga <0.1 Mo
<0.1 Sb <0.1 V <0.1 Bi <0.1 Gd <0.1 Na <0.5 Sc <0.1 W <0.1 Ca <0.5 Ge
<0.1 Nb <0.1 Se <0.1 Y <0.1 Cd <0.1 Hf <0.1 Nd <0.1 Sm <0.1 Yb <0.1 Ce
<0.1 Hg <0.2 Ni <0.2 Sn <0.1 Zn <0.5 Co <0.1 Ho <0.1 Pb <0.1 Sr <0.1 Zr
<0.1 Cr <0.1 In <0.1 Pd <0.1 Tb <0.1
Cs <0.1 K <0.2 Pr <0.1 Te <0.1

Dimethylformamide SpA

low metals

H255

1LT H255M
2½LT H255L
Dgr H:360D-226-312+332-319
P:201-210-302+352-305+351+338-
308+313



HCON(CH₃)₂ MW 73.09 BP 153.0°C d 0.95 CAS [68-12-2] Assay >99.9% Water
<0.05% Residue <0.0001%
Group 1 & 2 elements typically <0.1-<10 ppb
Group 3 to 12 (transition) elements typically <0.1-<5 ppb
Group 13, 14, 15 elements typically <0.1-<5 ppb
Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis (for organic analysis and chromatography see ROMIL Code H255 Dimethylformamide SpS)



Hydrochloric Acid SpA

H396

500ml H396P
1LT H396M
2½LT H396L
Dgr H:290-314-335
P:280c-301+330+331-305+351+338-
309+310



HCIMW 36.46 d 1.18CAS[7647-01-0] Assay 34-37% Colour <10 Hazen (APHA)
Ba, Be, Bi, Cd, Ce, Co, Cs, Dy, Er, Eu, Ga, Gd, Hf, Ho, In, La, Li, Lu, Mn, Mo, Nb,
Nd, Pb, Pr, Rb, Re, Rh, Ru, Sc, Sm, Sr, Tb, Te, Th, Ti, Tm, U, W, Y, Yb, Zr <0.1

ppb

each

As, Au, Cr, Cu, Hg, Mg, Ni, Sb, Sn, Ti, V <0.5 ppb each

Ag, Al, B, Ca, Fe, K, Na, Se, Zn <1 ppb each

Total P <0.01 ppm

Total S <0.3 ppm Free Cl2 <0.5 ppm

Br <10 ppm

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.1 Cs <0.1 La <0.1 Pt <0.5 Te <0.1 Al <0.5 Cu <0.1 Li <0.1 Rb <0.1
Th <0.1 As <0.1 Dy <0.1 Lu <0.1 Re <0.1 Ti <0.1 Au <0.1 Er <0.1 Mg <0.5
Rh <0.1 Ti <0.1 B <0.5 Eu <0.1 Mn <0.1 Ru <0.1 Tm <0.1 Ba <0.1 Fe <0.5
Mo <0.1 Sb <0.1 U <0.1 Be <0.1 Ga <0.1 Na <0.5 Sc <0.1 V <0.1 Bi <0.1
Gd <0.1 Nb <0.1 Se <0.1 W <0.1 Ca <0.5 Hf <0.1 Nd <0.1 Sm <0.1 Y <0.1
Cd <0.1 Hg <0.2 Ni <0.1 Sn <0.1 Yb <0.1 Ce <0.1 Ho <0.1 Pb <0.1 Sr <0.1
Zn <0.5 Co <0.1 In <0.1 Pd <0.5 Ta <0.5 Zr <0.1 Cr <0.1 K <0.1 Pr <0.1 Tb
<0.1

Hydro uoric Acid SpA

H405

500ml H405P
Dgr H:300+310+330-314
P:280-301+330+331-302+352-
304+340-305+351+338-
310+ROP010



HFIMW 20.01 d 1.16 CAS[7664-39-3] Assay 47-51%

Colour <10 Hazen (APHA)

Ba, Be, Bi, Cd, Ce, Co, Cs, Dy, Er, Eu, Ga, Gd, Ge, Hf, Ho, In, La, Li, Lu, Mn, Mo,
Nb, Nd, Pb, Pr, Rb, Re, Rh, Ru, Sc, Sm, Sr, Tb, Te, Th, Ti, Tm, U, V, Y, Yb, Zr <0.1

ppb each

Au, Pd, Pt, Sb <0.2 ppb each

Ag, As, Cu, Ni, Sn, W <0.5 ppb each

Al, B, Ca, Cr, Fe, Hg, K, Mg, Na, Se, Ti, Zn <1 ppb each

Total P <0.05 ppm

Total S <0.1 ppm

Cl <4 ppm

Sif6 <20 ppm

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

For treatment of HF burns, calcium gluconate gel (ROMIL Code PCG9V) is recommended. In order to provide emergency first aid, it should be kept wherever HF is handled or stored.

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.1 Cs <0.1 K <0.2 Pr <0.1 Tb <0.1
Al <0.5 Cu <0.1 La <0.1 Pt <0.2 Te <0.1
As <0.1 Dy <0.1 Li <0.1 Rb <0.1 Th <0.1
Au <0.1 Er <0.1 Lu <0.1 Re <0.1 Ti <0.5
B <0.5 Eu <0.1 Mg <0.2 Rn <0.1 Ti <0.1
Ba <0.1 Fe <0.5 Mn <0.1 Ru <0.1 Tm <0.1
Be <0.1 Ga <0.1 Mo <0.1 Sb <0.1 U <0.1
Bi <0.1 Gd <0.1 Na <0.5 Sc <0.1 V <0.1
Ca <0.5 Ge <0.1 Nb <0.1 Se <0.1 W <0.5
Cd <0.1 Hf <0.1 Nd <0.1 Sm <0.1 Y <0.1
Ce <0.1 Hg <0.05 Ni <0.1 Sn <0.1 Yb <0.1
Co <0.1 Ho <0.1 Pb <0.1 Sr <0.1 Zn <0.1
Cr <0.1 In <0.1 Pd <0.2 Ta <0.5 Zr <0.1

Hydrogen Peroxide SpA

H416

500ml H416P
Dgr H:302-318
P:280e-305+351+338-313



H2O2 MW 34.01d 1.10CAS[7722-04-1]

Assay ca. 30%

Colour <10 Hazen (APHA)

B, Be, Bi, Ce, Co, Cs, Dy, Er, Eu, Ga, Gd, Hf, Ho, Ir, La, Lu, Mn, Nd, Pd, Pt, Rb,
Re, Rh, Ru, Sc, Sm, Sr, Tb, Te, Th, Ti, Tm, U, Y, Yb <0.5 ppb each

Ba, Cu, Cr, Sb <1 ppb each

Ag, As, Cd, Ge, In, Li, Mg, Mo, Ni, Pb, Se, Sn, V, Zr <5 ppb each

Al, Ti, Zn <10 ppb each

Au, Ca, Fe, Hg, K, Na, Nb, Ta, W <50 ppb each

Cl, PO4 <0.5 ppm each

SO4, NO3 <1 ppm each

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

As a safety measure we fit bottles with a closure featuring a venting valve.

Typical values, elemental impurities at time of manufacture (ppb):

Ag 2 Cs <0.1 Ir <0.1 Pd <0.1 Tb <0.1 Al <0.1 Cu <0.1 K 15 Pt <0.1 Te <0.1
As <0.1 Dy <0.1 La <0.1 Rb <0.1 Th <0.1 Au <0.1 Er <0.1 Li <0.1 Re <0.1
Ti 4
B <0.1 Eu <0.1 Lu <0.1 Rh <0.1 Ti <0.1 Ba <0.1 Fe <6 Mg <0.1 Ru <0.1 Tm
<0.1 Ga <0.1 Mn <0.1 Sb <0.1 Bi <0.1 Gd <0.1 Mo <0.1 Sc
<0.1 V <0.1 Ca <0.1 Ge <0.1 Na 16 Se 2 W 1
Cd <0.1 Hf <0.1 Nb 0.2 Sm <0.1 Y <0.1 Ce <0.1 Hg 3 Nd <0.1 Sn <0.1 Yb
<0.1 Co <0.1 Ho <0.1 Ni <0.1 Sr <0.1 Zn 2
Cr <0.1 In <0.1 Pb <0.1 Ta 1 Zr 2



Methyl Alcohol (see Methanol)

4-Methylpentan-2-one (see Methyl iso-Butyl Ketone)

Methanol SpA

low metals

H413

500ml H413P
1LT H413M
2½LT H413L
Dgr H:225-301+311+331-370
P:210-280F-302+352-309+310-
403+235



(Methyl Alcohol)

CH3OH MW 32.04 BP 64.5°C d 0.79 CAS [67-56-1]
Assay >99.9% Water <0.05% Residue <0.0001%
Group 1 & 2 elements typically <0.5 ppb
Group 3 to 12 (transition) elements typically <0.5 ppb
Group 13, 14, 15 elements typically <0.5-
<50 ppb
Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis (for organic analysis and chromatography see ROMIL Code H410
Methanol SpS)

Typical values, elemental impurities at time of manufacture (ppb):

Ag	<0.5	Cd	<0.5	Mg	<0.5	Sn	<0.5
Al	<0.5	Co	<0.5	Mn	<0.5	Sr	<0.5
As	<0.5	Cr	<0.5	Mo	<0.5	Th	<0.5
B	29	Cu	<0.5	Na	<0.5	Ti	<0.5
Ba	<0.5	Fe	<0.5	Ni	<0.5	U	<0.5
Be	<0.5	Hg	<0.5	Pb	<0.5	V	<0.5
Bi	<0.5	K	<0.5	Sb	1	Zn	<0.5
Ca	<0.5	Li	<0.5	Se	<0.5	Zr	<0.5

Methyl iso-Butyl Ketone SpA

H439

500ml H439P
2½LT H439L
Dgr H:225-332-319-336-351-
EUH066
P:210-305+351+338-304+340



(4-Methylpentan-2-one)

(CH3)2CHCH2COCH3 MW 100.16 BP 117.4°C d 0.80 CAS [108-10-1]
Assay >99.8% Water <0.01% Residue <0.0001%
Group 1 & 2 elements typically <0.1-
<250 ppb
Group 3 to 12 (transition) elements typically <0.1-
<100 ppb
Group 13, 14, 15 elements typically <0.5-
<100 ppb
Peroxides (at time of manufacture) <0.0001% (<1 ppm)
Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis (for organic analysis and chromatography see ROMIL Code H446
Methyl iso-Butyl Ketone SpS)

Nitric Acid SpA

500ml H566P
1LT H566M
2½LT H566L
Dgr H:272-290-331-314-EUH071
P:260c-280c-301-330-331-
305+351+338-309+310



HNO3 MW 63.01 d 1.42CAS [7697-37-2] Assay 67-69% Colour <10 Hazen (APHA)
Ag, Au, Ba, Be, Bi, Ce, Cs, Dy, Er, Eu, Ga, Gd, Ge, Hf, Hg, Ho, In, La, Li, Lu, Mn,
Mo,
Nb, Nd, Pb, Pr, Rb, Re, Sc, Sm, Sr, Tb, Te, Th, Tl, Tm, U, W, Y, Yb, Zr <0.1 ppb
each

As, Cd, Co, Cu, Ni, Pd, Pt, Rh, Ru, Sb, Sn, Ti, V, Zn <0.5 ppb each
Al, B, Ca, Cr, Fe, K, Mg, Na, Se <1 ppb each

Total P <0.01 ppm

Cl <0.2 ppm

Total S <0.3 ppm

Store in dark.

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal Analysis, Ion Chromatography, Voltammetry

Elemental impurities specified at time of manufacture.

Concentrated Nitric Acid can decompose to nitrogen oxides (NOx) through action of heat or light resulting in a yellow colouration. However, this does not affect the performance of the acid with respect to trace metals or oxidising power.

Storage in a cool, dark place is recommended.

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.1 Cs <0.1 K <0.2 Pr <0.1 Tb <0.1
Al <0.5 Cu <0.1 La <0.1 Pt <0.1 Te <0.1
As <0.1 Dy <0.1 Li <0.1 Rb <0.1 Th <0.1
Au <0.1 Er <0.1 Lu <0.1 Re <0.1 Ti <0.1
B <0.5 Eu <0.1 Mg <0.2 Rh <0.1 Ti <0.1
Ba <0.1 Fe <0.5 Mn <0.1 Ru <0.1 Tm <0.1
Be <0.1 Ga <0.1 Mo <0.1 Sb <0.1 U <0.1
Bi <0.1 Gd <0.1 Na <0.2 Sc <0.1 V <0.1
Ca <0.5 Ge <0.1 Nb <0.1 Se <0.1 W <0.1
Cd <0.1 Hf <0.1 Nd <0.1 Sm <0.1 Y <0.1
Ce <0.1 Hg <0.02 Ni <0.1 Sn <0.1 Yb <0.1
Co <0.1 Ho <0.1 Pb <0.1 Sr <0.1 Zn <0.2
Cr <0.5 In <0.1 Pd <0.1 Ta <0.1 Zr <0.1



RIKI GLOBAL
performance through innovation

ROMIL-SpA™ Super Purity Acids and Reagents Specifications

Perchloric Acid SpA

less than 50% w/w

H675

500ml H675P
2½LT H675L
Dgr H:272-314
P:210-220-280c-301+330+331-
305+351+338-309+310



HClO₄ MW 100.46 d 1.40 CAS [7601-90-3] Assay 46-49% Colour <10 Hazen (APHA)

As, Au, Be, Bi, Ce, Co, Cs, Cu, Dy, Er, Eu, Ga, Gd, Ho, In, La, Li, Lu, Mo, Nd, Pd, Pr,

Pt, Rb, Rh, Sb, Sc, Sm, Sr, Tb, Te, Ti, Tm, U, V, Y, Yb, Zr <0.5 ppb each

Ag, Al, Ba, Ca, Cd, Fe, K, Mg, Mn, Na, Ni, Pb, Sn, Th, Ti, Zn <1 ppb each

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal

Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

Less hazardous acid concentration to facilitate international shipping.

Perchloric Acid SpA

H584

500ml H584P
2½LT H584L
Dgr H:271-290-314
P:210-221-280c-301+330+331-
305+351+338-309+310



HClO₄ MW 100.46 d1.66CAS[7601-90-3] Assay 65-71% Colour <10 Hazen (APHA)

As, Au, Be, Bi, Ce, Co, Cs, Cu, Dy, Er, Eu, Ga, Gd, Ho, In, La, Li, Lu, Mo, Nd, Pd, Pr,

Pt, Rb, Rh, Sb, Sc, Sm, Sr, Tb, Te, Ti, Tm, U, V, Y, Yb, Zr <0.5 ppb each

Ag, Al, Ba, Ca, Cd, Fe, K, Mg, Mn, Na, Ni, Pb, Sn, Th, Ti, Zn <1 ppb each

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal

Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.1 Cu <0.1 Li <0.1 Pt <0.5 Ti <0.5 Al <0.5 Dy <0.1 Lu <0.1 Rb <0.1 Ti

<0.1 As <0.1 Er <0.1 Mg <0.1 Rh <0.1 Tm <0.1 Au <0.1 Eu <0.1 Mn <0.1

Si <0.1 U <0.1 Ba <0.1 Fe <0.5 Mo <0.1 Sc <0.1 V <0.1 Be <0.1 Ga <0.1

Na <0.1 Sm <0.1 W <0.5 Bi <0.1 Gd <0.1 Nb <0.5 Sn <0.5 Y <0.1 Ca <0.5

Hf <0.5 Nd <0.1 Sr <0.1 Yb <0.1 Cd <0.1 Ho <0.1 Ni <0.5 Ta <0.5 Zn <0.5

Ce <0.1 In <0.1 Pb <0.1 Tb <0.1 Zr <0.1 Co <0.1 K <0.1 Pd <0.5 Te <0.1

Cs <0.1 La <0.1 Pr <0.1 Th <0.1

Sulphuric Acid SpA

H691

500ml H691P
1LT H691M
2½LT H691L
Dgr H:290-314
P:280c-301+330+331-305+351+338-
309+310



H₂SO₄ MW 98.07 d1.84CAS[7664-93-9] Assay 93-98% Colour <10 Hazen (APHA)

Ba, Be, Bi, Dy, Er, Eu, Ga, Gd, Hf, Ho, La, Lu, Nd, Pb, Pr, Sc, Sm, Tb, Te, Th, Tm,

U <0.1 ppb each

Au, Cd, Ce, Co, Cr, Cs, Cu, In, Li, Mn, Mo, Nb, Ni, Pt, Rb, Rh, Sr, Ti, V, W, Y, Yb,

Zr <0.5 ppb each

Ag, Al, As, Ca, Fe, Ge, Hg, K, Mg, Na, Sb, Sn, Ti, Zn <1 ppb each

Se <10 ppb

Total P <0.05 ppm

NO₃ <0.2 ppm

Cl <0.7 ppm

Substances reducing permanganate <20 ppm

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal

Analysis, Ion Chromatography, Voltammetry

Elemental impurities specific at time of manufacture.

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.1 Cu <0.1 La <0.1 Pt <0.5 Ti <0.5 Al <0.5 Dy <0.1 Li <0.1 Rb <0.5 Ti

<0.5 As <0.1 Er <0.1 Lu <0.1 Rh <0.5 Tm <0.1 Au <0.1 Eu <0.1 Mg <0.5 Sb

<1 U <0.1 Ba <0.1 Fe <0.5 Mn <0.5 Sc <0.1 V <0.1 Be <0.1 Ga <0.1 Mo

<0.1 Se <5 W <0.5 Bi <0.1 Gd <0.1 Na <0.5 Sm <0.1 Y <0.5 Ca <0.5 Ge

<0.1 Nb <0.5 Sn <0.1 Yb <0.5 Cd <0.1 Hf <0.1 Nd <0.1 Sr <0.1 Zn <0.2 Ce

<0.5 Hg <1 Ni <0.1 Ta <0.5 Zr <0.1 Co <0.1 Ho <0.1 Pb <0.1 Tb <0.1

Cr <0.1 In <0.5 Pd <0.5 Te <0.1

Cs <0.5 K <0.5 Pr <0.1 Th <0.1

Water SpA

H951

2½LT H951L

H₂0MW 18.02FP 0.0°C BP 100.0°C CAS [7732-18-5] Residue <0.0001% Resistivity (at time of manufacture) >18 MΩhm @ 25°C

Elemental impurities at time of manufacture:

Ag, Ba, Be, Bi, Cd, Hg, Mo, Pb, Sb <0.5 ppb each

Al, As, Co, Cr, Cu, Li, Mn, Ni, Sn, Sr, Th, Ti, U, V, Zr <1 ppb each

Fe, Mg, Zn <2 ppb each

B <10 ppb

Ca, K, Na <20 ppb each Equivalent to ASTM D1193 Type II

Application: Environment Analysis (eg, using AAS, ICP-OES, ICP-MS), Trace Metal

Analysis (for HPLC and organic trace analysis see ROMIL Code H950 Water SpS),

Ion Chromatography, Voltammetry

Typical values, elemental impurities at time of manufacture (ppb):

Ag <0.3 Ca <1 K <1 Pb <0.02

Al <0.2 Cd <0.01 Li <0.03 Sb <0.01

As <0.05

As <0.08 Co <0.2 Mg <0.2 Sn <0.03

Cr <0.2

B <6.5 Cr <0.03 Mn <0.02 Sr <0.02

Ge <0.01

Be <0.06 Cu <0.07 Mo <0.05 Th <0.02

Br <0.02 Fe <0.1 Na <1 Ti <0.05

Bi <0.02 Hg <0.01 Ni <0.05 U <0.02